

Test Report**No: SHCPCH160603538E****Date: Jun 20 2016**

Client name: Nantong City YOJO Medical Products Co.,LTD
Client address: Industrial Park,Sanxin Town,Haimen City,Jiangsu Province 226112 China

Sample name: PE WOUND PLASTER
Batch No./Date: 20160607
Manufacturer: Nantong City YOJO Medical Products Co.,LTD

The above sample and sample information were submitted and identified by the client.

SGS job No.: SHCPCH160603538
SGS reference No.: SHACPCH1612490601
Date of receipt: Jun 08 2016
Testing period: Jun 08 2016 ~ Jun 17 2016

TEST(S) REQUESTED:

Selected test(s) as requested by applicant:
1) Chemical test: Organic-tin compounds

TEST METHOD(S):

1) Organic-tin compounds: With reference to ISO 17353: 2004 with carbamate, analysis was performed by GC-MS.

TEST RESULT(S):

Please refer to next page

Remark: This test report is in English and maybe translated into other languages, The English version shall prevail.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and this document cannot be used for publicity without approval of the Company, not be allowed to copy testing report (except for copy of full text) without written approval.

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co.,Ltd



TEST RESULT(S):

Test Item(s)	Unit(s)	Test Method(s)	Test Result(s)	Method Detection Limit
Monobutyl tin (MBT)	mg/kg	1)	Not detected	0.02
Dibutyl tin (DBT)	mg/kg		Not detected	0.02
Tributyl tin (TBT)	mg/kg		Not detected	0.02
Monoocetyl tin (MOT)	mg/kg		Not detected	0.02
Tetrabutyl tin (TTBT)	mg/kg		Not detected	0.02
Diocetyl tin (DOT)	mg/kg		Not detected	0.02
Triphenyl tin (TPhT)	mg/kg		Not detected	0.02
Tricyclohexyl tin (TCyT)	mg/kg		Not detected	0.02
Tributyl Tin Oxide (TBTO) *	mg/kg		Not detected	0.02

Notes : *Calculated concentration of TBTO is based on the identified TBT.

SAMPLE DESCRIPTION: Wound plaster



*** End of Report***

